

amc technical brief

Analytical Methods Committee

No. 13 Sep 2003

© Royal Society of Chemistry 2003

Unfortunately we are still somewhat lax about terminology. How often do we say .0028... 4 TMC/P MCID 7 BD5 Tsi,oTge23 Tt0still some

Note – The term accuracy, when applied to a set of test results, involves a combination of random components and a common systematic error or bias component. [ISO 3534: 3.11]

AMC comments

- Accuracy is essentially absence of error. A result of higher accuracy has a smaller error.
- *Accuracy* should not be used in contrast to *precision* – a result is unlikely to be accurate if the results in general are not precise.
- Notice also that, strictly, *accuracy* applies to results and not more general entities such as analytical methods, laboratories or

- ISO 5725 additionally discusses *intermediate measures of precision*, and provides a notation for conditions in which time, calibration, operator and equipment are varied.
- With the help of these specified conditions and some other common terms, it becomes straightforward to describe estimates of precision. Some important examples follow.

Repeatability Precision estimated under repeatability conditions. [ISO 3534: 3.15]

Reproducibility Precision under reproducibility conditions. [ISO 3534: 3.20]

Run-to-run precision